



**DATA VALIDATION CHECKLIST – STAGE 2A
EPA REGION 8 START CONTRACT**

Site Name	Telluride Valley Floor Removal Action	TO/TOLIN No.	71-2107-06
Document Tracking No.	NA	Technical Reviewer (signature and date)	 August 24, 2021
Data Reviewer (signature and date)	 8/24/2021	Laboratory	Eurofins TestAmerica/Phoenix, AZ
Laboratory Report No.	550-169030-1		
Analyses	Arsenic and lead by NIOSH 7303		
Samples and Matrix	Seven air samples, including one field duplicate sample and one field blank		
Field Duplicate Pairs	VF003081221/VF003081221D		
Field Blanks	VF001081221FB		

INTRODUCTION

This checklist summarizes the Stage 2A validation performed on the subject laboratory report, in accordance with the U.S. Environmental Protection Agency (EPA) *Guidance for Labeling Externally Validated Laboratory Analytical Data for Superfund Use* (January 2009). Analytical data were evaluated in general accordance with *START V Programmatic Quality Assurance Project Plan* (October 2020) and the EPA *National Functional Guidelines (NFG) for Inorganic Superfund Methods Data Review* (January 2017).

OVERALL EVALUATION

No rejection or qualification of data was required for this work order. The data can be used as reported by the laboratory.

Data completeness:

Within Criteria	Exceedance/Notes
Y	

Sample preservation, receipt, and holding times:

Within Criteria	Exceedance/Notes
N	The EDD erroneously reported sample VF001080921 as VF0010809321. This sample name was manually corrected in the attachment.

**DATA VALIDATION CHECKLIST – STAGE 2A
EPA REGION 8 START CONTRACT**

Method blanks:

Within Criteria	Exceedance/Notes
Y	

Field blanks:

Within Criteria	Exceedance/Notes
Y	

System monitoring compounds (surrogates and labeled compounds):

Within Criteria	Exceedance/Notes
NA	

MS/MSD:

Within Criteria	Exceedance/Notes
NA	

Laboratory duplicates:

Within Criteria	Exceedance/Notes
NA	

Field duplicates:

Within Criteria	Exceedance/Notes
Y	

**DATA VALIDATION CHECKLIST – STAGE 2A
EPA REGION 8 START CONTRACT**

LCSs/LCSDs:

Within Criteria	Exceedance/Notes
Y	

Sample dilutions:

Within Criteria	Exceedance/Notes
NA	

Re-extraction and reanalysis:

Within Criteria	Exceedance/Notes
NA	

MDLs/RLs:

Within Criteria	Exceedance/Notes
Y	

Tentatively identified compounds:

Within Criteria	Exceedance/Notes
NA	

Other [specify]:

Within Criteria	Exceedance/Notes
NA	

DATA VALIDATION CHECKLIST – STAGE 2A EPA REGION 8 START CONTRACT

Overall Qualifications:

See results summary pages attached for changes to the laboratory qualifiers based upon this validation. The following is a list of qualifiers and definitions that may be used for the validation of this data package:

J	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample.
J+	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample and may be biased high.
J-	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample and may be biased low.
NJ	The analysis indicates the presence of an analyte that has been “tentatively identified” and the associated value is the approximate concentration of the analyte in the sample.
R	The sample result is rejected as unusable due to serious deficiencies in one or more quality control criteria. The analyte may or may not be present in the sample.
U	The analyte was analyzed for, but was not detected at or above the associated value (reporting limit).
UJ	The analyte was analyzed for, but was not detected at or above the associated value (reporting limit), which is considered approximate due to deficiencies in one or more quality control criteria.

TELLURIDE VALLEY FLOOR REMOVAL ACTION AIR ANALYTICAL RESULTS SUMMARY
EUROFINS TESTAMERICA REPORT NO. 550-169030-1

Samp_No	Analysis	Analyte	Lab Result	Lab Qual	RL	Units	VAL_Result	VAL_Qual
VF001080921	NIOSH 7303	Arsenic	0.0739	U		0.0739 ug/m3	0.0739	U
VF001080921	NIOSH 7303	Lead	0.0369	U		0.0369 ug/m3	0.0369	U
VF001080921	NIOSH 7303	Arsenic	0.5	U		0.5 ug/Sample	0.5	U
VF001080921	NIOSH 7303	Lead	0.25	U		0.25 ug/Sample	0.25	U
VF001081221	NIOSH 7303	Arsenic	0.0674	U		0.0674 ug/m3	0.0674	U
VF001081221	NIOSH 7303	Lead	0.0337	U		0.0337 ug/m3	0.0337	U
VF001081221	NIOSH 7303	Arsenic	0.5	U		0.5 ug/Sample	0.5	U
VF001081221	NIOSH 7303	Lead	0.25	U		0.25 ug/Sample	0.25	U
VF001081221FB	NIOSH 7303	Arsenic	0.5	U		0.5 ug/Sample	0.5	U
VF001081221FB	NIOSH 7303	Lead	0.25	U		0.25 ug/Sample	0.25	U
VF002081221	NIOSH 7303	Arsenic	0.181	U		0.181 ug/m3	0.181	U
VF002081221	NIOSH 7303	Lead	0.0903	U		0.0903 ug/m3	0.0903	U
VF002081221	NIOSH 7303	Arsenic	0.5	U		0.5 ug/Sample	0.5	U
VF002081221	NIOSH 7303	Lead	0.25	U		0.25 ug/Sample	0.25	U
VF003080921	NIOSH 7303	Arsenic	0.161	U		0.161 ug/m3	0.161	U
VF003080921	NIOSH 7303	Lead	0.0807	U		0.0807 ug/m3	0.0807	U
VF003080921	NIOSH 7303	Arsenic	0.5	U		0.5 ug/Sample	0.5	U
VF003080921	NIOSH 7303	Lead	0.25	U		0.25 ug/Sample	0.25	U
VF003081221	NIOSH 7303	Arsenic	0.164	U		0.164 ug/m3	0.164	U
VF003081221	NIOSH 7303	Lead	0.0818	U		0.0818 ug/m3	0.0818	U
VF003081221	NIOSH 7303	Arsenic	0.5	U		0.5 ug/Sample	0.5	U
VF003081221	NIOSH 7303	Lead	0.25	U		0.25 ug/Sample	0.25	U
VF003081221D	NIOSH 7303	Arsenic	0.164	U		0.164 ug/m3	0.164	U
VF003081221D	NIOSH 7303	Lead	0.0818	U		0.0818 ug/m3	0.0818	U
VF003081221D	NIOSH 7303	Arsenic	0.5	U		0.5 ug/Sample	0.5	U
VF003081221D	NIOSH 7303	Lead	0.25	U		0.25 ug/Sample	0.25	U